

Certificate of Analysis

Product **Petroleum Ether 40 - 60** **Batch No.** : **21 05 0181**
Grade **AR.** **Manufacturing date** : **21/05/2021**
Code No. **AR1147** **Best before** : **05/2025**

| Items | Descriptions | Specifications | | Test Results |
|--------------|------------------------|-----------------------|------|---------------------|
| 1 | Water (by Coulometry) | 0.02% | max. | 0.007% |
| 2 | Color (APHA) | 10 | max. | < 10 |
| 3 | Acidity (mEq./g.) | 0.0005 | max. | < 0.0005 |
| 4 | Residue on Evaporation | 0.001% | max. | < 0.001% |
| 5 | Sulfur Compounds (S) | 0.002% | max. | < 0.002% |
| 6 | Boiling range (°C) | 40 - 60 | | 40 - 60 |

Meet A.C.S. Specifications.

This is to certify the above batch has been tested and complies with this specification.

Ladaphan.

Chemist

Achan.

Technical Manager