RCI Labscan

RCI Labscan Limited

Certificate of Analysis

Product	**BUFFER SOLUTION pH 10.0 (Blue)**	**Batch No.**	**:**	**21 05 0323**
Code No.	**GN1023**	**Manufacturing date**	**:**	**31/05/2021**
		Best before	**:**	**05/2023**

Items	Test	Specifications	Test Results
1	Appearance	Clear, blue solution	Passes test
2	pH at 25°C	pH 10.00 ± 0.02	pH 10.00

Traceable to NIST SRMs : 191d

This is to certify the above batch has been tested and complies with this specification.

Deviations of pH (△pH) at various temperatures :

	△pH
5°C	+ 0.20
10°C	+ 0.11
15°C	+ 0.07
20°C	+ 0.04
25°C	± 0
30°C	- 0.09
35°C	- 0.12
40°C	- 0.15
50°C	- 0.20

Ladaphan

Chemist

Technical Manager