RCI Labscan

RCI Labscan Limited

Certificate of Analysis

Product	**Acetonitrile**	**Batch No.**	:	**21 06 0105**
Grade	**AR.**	**Manufacturing date**	:	**08/06/2021**
Code No.	**AR1005**	**Best before**	:	**06/2025**

Items	Descriptions	Specifications		Test Results
1	Assay (by GC.)	99.7%	min.	99.9%
2	Appearance	Clear		Clear
3	Color (APHA)	10	max.	< 10
4	Water (by Coulometry)	0.05%	max.	0.012%
5	Acidity (mEq./g.)	0.0005	max.	< 0.0005
6	Alkalinity (mEq./g.)	0.0006	max.	< 0.0006
7	Residue on Evaporation	0.001%	max.	< 0.001%
8	Propionitrile	0.2%	max.	< 0.2%

Meet A.C.S. Specifications.

This is to certify the above batch has been tested and complies with this specification.

Ladaphan.

Chemist

Technical Manager