

Certificate of Analysis

Product	Methanol	Batch No.	: 21 06 0314
Grade	Super Gradient for HPLC	Manufacturing date	: 14/06/2021
Code No.	SG1115	Best before	: 06/2025

Items	Descriptions	Specifications	Test Results
1	Assay (by GC.)	99.9% min.	99.9%
2	Identity (IR)	Passes test	Passes test
3	Color (APHA)	10 max.	< 10
4	Water (by Coulometry)	0.02% max.	0.016%
5	Acidity (mEq./g.)	0.0003 max.	< 0.0003
6	Alkalinity (mEq./g.)	0.0002 max.	< 0.0002
7	Residue on Evaporation	0.0002% max.	< 0.0002%
8	Acetone (GC.)	0.001% max.	< 0.001%
9	UV Transmission Levels (%T)		
	250 nm	99% min.	99.9%
	240 nm	98% min.	99.4%
	230 nm	90% min.	98.0%
	220 nm	75% min.	94.9%
	210 nm	65% min.	78.3%
10	Gradient Specification		
	at 230 nm	2.0 mAU max.	< 2.0 mAU
	at 235 nm	2.0 mAU max.	< 2.0 mAU
	at 254 nm	1.0 mAU max.	< 1.0 mAU
11	Fluorescence (as quinine)		
	at 254 nm	1 ppb max.	< 1 ppb
	at 365 nm	0.5 ppb max.	< 0.5 ppb

Conforms to Reag. Ph.Eur

Suitable for HPLC, UPLC / UHPLC / Ultra HPLC - instruments.

Product passed through 0.2 micron final filter.

This is to certify the above batch has been tested and complies with this specification.

Ladaphan

Chemist

Achan

Technical Manager