RCI Labscan Limited

Certificate of Analysis

Product	**BUFFER SOLUTION pH 4.0 (Red)**	**Batch No.**	**:**	**21 06 0303**
Code No.	**GN1017**	**Manufacturing date**	**:**	**23/06/2021**
		Best before	**:**	**06/2023**

Items	Test	Specifications	Test Results
1	Appearance	Clear, red solution	Passes test
2	pH at 25°C	pH 4.00 ± 0.02	pH 4.00

Traceable to NIST SRMs : 185i

This is to certify the above batch has been tested and complies with this specification.

Deviations of pH (△pH) at various temperatures :

	△pH
5°C	- 0.01
10°C	- 0.01
15°C	- 0.01
20°C	- 0.01
25°C	± 0
30°C	+ 0.01
35°C	+ 0.02
40°C	+ 0.02
50°C	+ 0.03

Ladaphan

Chemist

Technical Manager